

IMPORTANT NOTICE

REGARDING LUCIRA COVID-19 TEST KIT EXPIRY DATE

Thank you for your order. We want to provide you updated information about the expiration date on your Lucira COVID-19 Test Kits.

The shelf life for Lucira COVID-19 Test Kits has been extended to 12 months. For COVID-19 tests, the FDA only allows manufacturers to label products based on real-time stability studies. As of September 1st, 2021, Lucira Health has completed the real-time stability testing to support a 12-month shelf life.

You will find an updated expiration date stickered on the bottom of each test carton enclosed in this shipment. Some of the internal pouches enclosed within each test kit may have been printed with a 9-month expiry label that has now been extended an additional 3 months. The expiration date on the test carton reflects the updated current expiration date for all internal test kit components.

If you have previously ordered test kits from Lucira, please note that those expiration dates have also been extended. The table below summarizes the newly extended expiration dates:

Printed Expiration Date	New Expiration Date
June 2021	September 2021
July 2021	October 2021
August 2021	November 2021
September 2021	December 2021
October 2021	January 2022
November 2021	February 2022

Printed Expiration Date	New Expiration Date
December 2021	March 2022
January 2022	April 2022
February 2022	May 2022
March 2022	June 2022
April 2022	July 2022
May 2022	August 2022

Please note that we also have accelerated aging studies underway that presently show product stability at 18 months. We will be continuing our real-time stability studies until we reach an 18-month shelf life on our product.

If you have any questions, please call us at 1-888-582-4724 or email us at orders@lucirahealth.com.

Thank you,

The Lucira Team